

# Opinion: Should high resolution differential mobility analyzers be used in mainstream aerosol studies?

Juan Fernandez de la Mora

Department of Mechanical Engineering & Materials Science, Yale University, New Haven, CT 06520,

5 USA

Correspondence to: Juan Fernandez de la Mora (juan.delamora@yale.edu)

**Abstract.** High Resolution Differential Mobility Analyzers (HRDMAs) are defined here as plain DMAs maintaining a steady flow over an unusually broad range of sheath gas flow rates  $Q$ . HRDMAs, first developed by Georg Reischl's group (Winklmayr et al, 1991), have existed for a long time. However, they have not been widely adopted, except in the size range below 10 nm, often in new particle formation studies (Kangasluoma et al. 2020 and Ozon et al. 2021). Here we question the notion that HRDMAs are necessarily complex, bulky and expensive machines, mainly of interest in exotic applications outside mainstream aerosol research. Rather, many studies central to aerosol research could be carried out with HRDMAs with considerable advantage in size range, resolution, sensitivity and measurement speed. DMA manufacturers will hopefully take the challenge of developing commercial HRDMAs of complexity and cost comparable to those of today's commercial instruments, adapted for broad use by aerosol scientists, though with greatly improved flexibility and performance. Some of the technical challenges that still need to be overcome are also discussed, such as the development of high flow condensation counter detectors, and the control of high sample and sheath gas flow rates.

Formatted: Space Before: Auto, After: Auto

Deleted: and his students

Deleted: , yet

## 1. Introduction

Flagan (1998) has described the long and complex evolution of electrical aerosol measurement methods, culminating in the modern development of the differential mobility analyzer (DMA) by Liu and Pui (1974) and Knutson and Whitby (1975). DMAs have subsequently become irreplaceable instruments, extensively used in studies involving submicron particles. Here we shall focus on high resolution DMAs (HRDMAs), defined as DMAs operating steadily at unusually high flow rate of sheath gas (HFDMA). One possible quantitative definition is HRDMA=HFDMA=DMA operating steadily at  $Q > 100$  L/min. It is nevertheless preferable to define HRDMAs more qualitatively as DMAs including special features enabling an extended  $Q$  range. Likewise, we shall denote as plain DMAs those having no such special features.

A distinction must be made in this respect between subcritical, critical and supercritical DMAs, based on the fact that laminar flow in a tube, or in a channel between two plates (or two concentric tubes) typically becomes unstable at Reynolds numbers (Re) in the range of 2000 (depending on geometry). Supercritical DMAs may roughly be defined as those keeping the flow laminar well above  $Re=2000$  by

carefully laminarizing the sheath flow, and by also avoiding any slight discontinuity on the wetted walls. Critical DMAs could be similarly defined as those keeping the flow laminar up to  $Re=2000$ , but not much beyond. Subcritical DMAs would then be those including unusual features forcing flow instabilities well below  $Re=2000$ . This distinction is not irrelevant because laminar DMA operation at  $Re$  well above 100,000 has been demonstrated many times, while most commercial DMAs operate well below  $Re=2000$ . Accordingly, because critical DMAs would represent a substantial improvement over the usual current situation, we will treat them as HRDMAs for the present discussion.

Only a minor fraction of the large bibliography on DMAs has centered on applications benefitting from a relatively high resolution, mostly focusing on particles smaller than 10 nm. This has created an impression that the main utility of HRDMAs is in the low *nano* corner of aerosol research, rarely well resolved by more conventional DMAs. The question examined in this *opinion* is whether it would make sense to use future HRDMAs in situations presently handled by *plain* DMAs. We anticipate that the

answer will be YES, primarily due to considerable benefits in terms of size range, resolution and sensitivity.

At first sight it would appear that HRDMAs are unlikely to play a role in mainstream aerosol research for a variety of reasons. There is first the perception that HRDMAs are complex, heavy, and expensive machines, requiring large pumps, cooling systems, long diffusers, wide inlets, etc. For instance,


Kangasluoma et al. 2017 note that “However, these supercritical DMAs were mainly used in laboratory calibrations in previous studies because of their high flow rates of several hundred or thousand liters per minute and the corresponding high maintaining expenses”. Similarly, we read in Cai et al. 2017 (2010) that “However, these DMAs are designed to operate at high sheath flowrate (typically several hundreds to thousands lpm) rather than matching the flowrate required by the DEG-UCPC. In addition, large sheath flow recirculation system (including high volume filters and air mover) are required to operate those DMAs, making it troublesome for field measurement”.

These remarks must be taken most seriously, as they originate from groups working on new particle formation, which are well acquainted with the characteristics of HRDMAs. More fundamentally, many aerosol studies cover a vast size range, rarely including narrowly defined features demanding high resolution. Furthermore, if this extensive size range were probed at high resolution, it would apparently take a long time to do so, with each narrow size range individually examined containing too little signal. These theoretical reasons seem to be confirmed by the observation that the vast majority of DMA users have relied on long established commercial instruments operating at resolving powers typically of 10 or less. DMA resolution is defined here as the inverse of the relative full width at half maximum (*FWHM*)

of the transfer function.

75

80



85 Figure 1: Sketch of a cylindrical DMA with grounded outer electrode of radius  $R_2$  and inner electrode of radius  $R_1$  held at potential  $V$ , with a volumetric flow rate  $Q$  of sheath gas passing symmetrically between both electrodes

To guide the process of re-examining these various apparently persuasive notions against the wider use of HRDMAs, let us consider the main operating DMA parameters: the flow rates of polydisperse and monodisperse aerosol,  $q$  (taken for simplicity to be identical), the flow rate  $Q$  of sheath gas, the classification voltage  $V$  (Figure 1), and the mean electrical mobility  $Z$  of the classified particles. For planar and axisymmetric instruments, these quantities are simply connected by equation (1) through a single parameter  $k$ , fixed by the instrument geometry:

$$Z = kQ/V. \quad (1)$$

The relative width of the transfer function,  $FWHM$ , also depends on  $q$ ,  $Q$ ,  $V$  and  $Z$ , mainly through  $q/Q$ , and  $V$  (Knutson and Whitby, 1975; Flagan, 1999)

$$FWHM = q/Q \text{ for non-diffusive particles,} \quad (2)$$

$$FWHM^2 \sim V/(k_B T) \text{ for diffusive particles when } q \ll Q, \quad (3)$$

where  $k_B T$  is the thermal energy.

Typically,  $k$  and  $Z$  are externally imposed, with little room to maneuver (though  $k$  may be scanned in DMAs of variable geometry: Bezantakos et al., 2016; Lee et al., 2020; Perez Lorenzo et al., 2020a). The control parameters most readily available to the user are accordingly  $Q$  and  $V$ . The voltage is in most DMAs widely variable over an extended range, from a few volts up to the maximum value prior to electrical breakdown, typically between 5 and 15 kV, depending on geometry. The advantage of this ample  $V$  range is evident from Equation (1), as it enables spanning a comparable range in particle mobility. In contrast,  $Q$  is most often varied over a much narrower range, from about 1 L/min to 40 L/min. Why this limited  $Q$  range is at first sight puzzling, as, in view of (1), the accessible  $Z$  range depends as much on  $V$  as on  $Q$ . One's puzzlement increases further when noting that the resolving power also depends as strongly on  $V$  as on  $Q$ , since both are coupled through (1). A DMA with a narrow  $Q$  range necessarily offers far less operational flexibility than another with a wider  $Q$  range. We shall see that a limited  $Q$  range implies a limited performance not only in terms of size range and resolving power, but also in sensitivity, and speed of response.

## 2 DMA characteristics

### 2.1 Size range, resolving power, sensitivity and optimal scan.

According to Eq. (1), a mobility spectrum may be acquired as well by scanning over  $Q$  as over  $V$ .  
115 Suppose one needs to cover the mobility range  $Z_{max} \leq Z \leq Z_{min}$ . Given that the diffusion-limited resolving power increases as  $V^{1/2}$ , [Equation (3)] the optimal scan would start with the largest particle,  $Z_{min} = kQ_{min}/V_{max}$ , and evolve over increasing  $Q$  values, with  $V$  fixed at  $V_{max}$ . An additional advantage of this  $Q$  scan over conventional  $V$  scans (besides its minimization of diffusive broadening) is that one may increase  $q$  proportionally to  $Q$  to improve the sensitivity without resolution loss. Once the limit  $Q = Q_{max}$  is reached, the mobility spectrum may be extended by scanning over decreasing  $V$  at fixed  $Q = Q_{max}$ . The range  $Z_{max}/Z_{min}$  in this scan is  $Q_{max}V_{max}/(Q_{min}V_{min})$ . This range applies equally to *plain* and high resolution DMAs. The difference is that the former may typically vary  $Q$  from 2 to 25 L/min, while the later may reach well beyond 1000 L/min. Some level of  $Q$  extension (not necessarily up to thousands of L/min) is well known to be essential to analyze 1 nm particles without excessive diffusive broadening.

### 125 2.2 Response time.

Another important advantage of HRDMAs relates to permissible scan speed (Fernandez de la Mora et al., 2017a). High  $Q$  implies a reduced response time, but fast measurements do not strictly require high flow rates. High resolution is even more useful, as it implies that particles of a given  $Z$  are classified over a narrow range of  $V/Q$  values. As a result, they exit the DMA at almost the same time during a mobility scan. If the scan is fast, the spectrum will tend to be distorted both in *plain* DMAs and HRDMAs. The reason is that, at the time the DMA voltage is  $V(t)$ , the detector senses a signal  $I(t-\Delta t)$  corresponding to the DMA voltage (or flow rate) applied a certain time delay  $\Delta t$  earlier. Nevertheless, the mobility spectra in HRDMAs are undistorted (other than through this uniform time shift), whence the original size spectrum may be recovered via a simple translation of the voltage in time. In contrast, 135 in *plain* DMAs, particles of a given mobility are classified over a wider range of  $V/Q$  (scan times), resulting in peak shape distortions that are not simple time delays. Recovery of the undistorted signal is in this case far from trivial, forcing considerably slower scans than achievable in HRDMAs. Greater speed of measurement is a characteristic of universal interest in all situations where the signal is strong enough. This has led to the recent commercial development of a variety of condensation particle counter 140 detectors (CPCs) with relatively fast responses. Why would DMA manufacturers not take advantage of this notable advance by developing HRDMAs generating undistorted size spectra with response times comparable or better than those of contemporary fast CPCs?

### 2.3 Presumed irrelevance of HRDMAs.

Let us provisionally assume that high resolution is not necessary in most aerosol studies spanning a wide size range, on the grounds that no narrow features exist in typical size spectra. Let us further 145 accept that not enough signal is available over narrow size regions, and that most aerosol studies do not need to cover the size range below 5-10 nm. A HRDMA would still be substantially more useful than a plain DMA. Indeed, thanks to its broad  $Q$  range, its resolving power and size range can be controlled far

more widely than in plain DMAs. Complete freedom to run HRDMAs at limited resolution, when  
150 desired, naturally hinges on the ability to increase the aerosol sample flow  $q$  proportionally to  $Q$ . This  
would simultaneously greatly increase sensitivity, largely controlled by  $q$ . Increasing  $q$  is  
straightforward with existing electrometer detectors. However, Susanne V. Hering (private  
communication) has perceptively pointed out that high flow CPCs do not presently exist. In order to  
flexibly increase the sensitivity and control the resolution of HFDMAs, one would need to develop  
155 them together with high flow CPCs. This double challenge may seem non-trivial, but success in it  
would have a large impact on aerosol research.

We have so far considered studies of particles larger than 5-10 nm at moderate resolving power. There  
are nevertheless situations of clear interest calling for the classification and detection of ultrafine  
particles with resolving powers higher than available with plain DMAs. An example is the formation of  
160 new atmospheric particles and their subsequent growth (Kong, et al. 2021). If the nucleation event is  
brief and the growth period long, as often happens, the new particles will tend to have narrow size  
distributions. A plain DMA will not determine as fast, sensitively, and accurately either the particle size  
or the growth rate. It will not even see a narrow size distribution if it exist. It is not widely realized that  
the sensitivity with which one can detect narrowly defined size distributions increases rather than  
165 decreasing at increasing resolving power. We observe this all the time in the study of viruses, with  
peaks either narrow and isolated, or wider and partially buried within a large background, at high and  
low flow rates, respectively (See Figures 2 and 3 of Fernandez de la Mora et al. 2021). The reason is  
that the signal is concentrated over a narrow mobility range, while the noise is spread continuously over  
a much wider domain. Capturing the whole signal over a narrow range of mobilities therefore reduces  
170 greatly the noise **but not the signal**. In contrast, capturing the particles over a wider mobility window  
does not increase the signal, yet augments the noise. The ideal resolving power from the strict point of  
view of signal/noise is accordingly dependent on circumstances. A flexible instrument where the  
resolving power and the sensitivity may be tuned as required by these circumstances is evidently better  
suited for this and other comparably demanding applications. Likewise, a flexible single instrument able  
175 to cover the nanorange (Perez-Lorenzo et al. 2021) as well as 200-300 nm particles (Fernandez de la  
Mora et al. 2023; Fernandez de la Mora and Papanu, 2023) is far more convenient than investing in two  
different instruments, one for each of these two ranges. Not to mention the issue of matching in a single  
size spectrum the outputs of two devices with different characteristics.

#### 2.4 Widening the $Q$ range: is it so hard?

180 The simple criterion adopted here based on flow rate of sheath gas appears to facilitate the classification  
of the numerous commercial DMAs in the market. Nevertheless, even when the manufacturer indicates  
a modest maximal or operating  $Q$ , the instrument may in fact accept much larger flows. For instance, in  
their first detailed description of Reischl's short 1/40 DMA (nominally classifying particles from 1 to 40  
nm in diameter, but really going up to 150 nm), Winklmayr et al. (1991) indicated an operational flow  
185 rate  $Q=28$  L/min. However, Rosell et al. (1996) found that they could draw  $Q$  values beyond 300 L/min  
with hardly any changes, and even beyond 800 L/min by adding two more exhaust lines to the single  
original exhaust in the sheath gas manifold (de Juan et al. 1998). Furthermore, their shorter version of

this 1/40 DMA maintained the flow laminar over most of this considerable  $Q$  range. Reischl's short 1/40 DMA therefore qualifies *de facto* as the first HRDMA.

190 The first surprising feature we discovered in our study of TSI's 3071 DMA was that the Reynolds number in its annular classification region was only a few hundreds at the highest recommended flow rates of 20-40 L/min, at which some flow instability was already present. These are the flow rates typical of other commercial DMAs, to which this puzzling situation may also apply. One would certainly expect a serious deterioration of the performance at increasing Reynolds numbers due to  
195 turbulent transition. But this transition is not supposed to take place until Reynolds numbers ( $Re$ ) well above 1000. And even above that critical value, turbulent transition takes a considerable length to develop, especially when the inlet flow has been carefully laminarized, or when the working section is slightly converging. As theoretically expected, the unnaturally precocious flow instability in TSI's 3071  
200 DMA could be removed by avoiding two types of aerodynamic problems: (i) steps following immediately after the inner and outer radii of the laminarization screen, and (ii) unstable regions with decelerating boundary layers in the sheath gas inlet (Eichler et al. 1998; Fernandez de la Mora et al.  
205 2017b). Once these aerodynamic **extravagances** were cleared, widening a few downstream features offering excessive flow resistance enabled reaching  $Q=100$  L/min without any signs of flow instability. However, although the resolution increased considerably, it remained well below the theoretical value dictated by Brownian diffusion and by the finite value of  $q/Q$ . This latter result suggested that flow perturbations injected at the laminarization screen are a real problem in strictly cylindrical DMAs. Indeed, since the flow cross section is the same in the laminarization screens and in the classification region, the constriction created by the screens accelerates locally the flow into a multitude of jets, whose decay is by no means immediate. Accordingly, the famous 3071 DMA and various generations  
210 of successors still limited to flow rates below 25-40 L/min at TSI and elsewhere, must presently be classified as *plain* DMAs. Nevertheless, based on the precedents just discussed, some among these *plain* instruments could possibly approach and even reach the HRDMA category.

Deleted: blunders

### 2.5 Reischl's inlet trumpet and its minimization.

The screen problem just described for strictly cylindrical DMAs must have been known to Georg  
215 Reischl when he developed the first HRDMA featuring a trumpet-shaped sheath flow inlet, such that the cross section of the laminarization screens was substantially wider than that in the analyzing region. The trumpet included in Reischl's 10/40 DMA deserves some comment, as it was not discussed in any of Reischl's published articles, and does not even appear in the *schematic* in Figure 1 of Winklmayr et al. (1991). This trumpet-less published *schematic* was apparently used by others in the development of  
220 various clones, widely circulated in European laboratories. Some at least among these copies did not enjoy the extended  $Q$  characteristics of the original design, and were the cause of some confusion. For instance, Rosell et al. (1996) include the following footnote. "... a preliminary test of a shortened Reischl DMA was performed .... However, for reasons never fully understood, that short DMA did  
225 neither yield the predicted resolution, nor did it operate properly at flow rates in the range of 80 L/min or above." The mystery noted was simply the absence of that inlet trumpet in the cloned prototype used, as clarified years later in a private conversation with George Reischl (who stressed the distinction  
between a *schematic* and a *drawing*, as further clarified in the historically relevant discussions to this

230 | [opinion by J. Rosell and Gerhard Steiner](#)). Fortunately, the successful study by Rosell et al. (1996) had  
the benefit of Reischl's original drawings, including the inlet trumpet, liberally shared by their inventor  
with colleagues who requested them. Therefore, this inlet contraction must be taken to be an essential  
element in HRDMAs, at least until an alternative approach is demonstrated. An inlet trumpet was  
certainly part of all the successful DMAs developed at Yale. It is featured not only in prototypes and  
fabrication drawings of Reischl's 1/40 DMA, but also in publications describing his later high- $Q$   
235 | designs (Steiner et al., 2010; Keck et al. 2008), some of which were commercialized by Grimm [Aerosol  
Technik GmbH](#). These more recent Reischl DMAs have been successfully tested at relatively high flow  
rates and do undoubtedly qualify as HRDMAs. The one Reischl model possibly belonging to the *plain  
category* is the long 10/1000 DMA, shown schematically in Figure 2 of Winklmayr et al. (1991). This  
DMA features a widening rather than a converging section following the laminarization region, and  
240 | operates nominally at  $Q=12$  L/min. If the published *schematic* was faithful to the actual design, it is  
improbable that the 10/1000 DMA would have sustained a steady sheath flow at substantially higher  
 $Q$ s.

245 | [On the assumption that an inlet trumpet is essential](#), an important practical issue is to determine the  
smallest cross section ratio  $A_s/A_w$  between the open screen area and the [working](#) section required by a  
HRDMA. I am not aware of any systematic study aimed at minimizing  $A_s/A_w$ , though our experience  
with TSI's 3071 DMA suggests that this area ratio must exceed unity. This does not necessarily mean  
that HRDMAs must be heavy and bulky. Three examples at least of hand-held HRDMAs have been  
described: Reischl's 10/40 (Winklmayr et al., 1991), the Half-Mini (Fernandez de la Mora, 2017) and  
the earliest version of the Perez DMA family (Perez-Lorenzo et al., 2020b). Worthy of note is the fact  
250 | that Martinez-Lozano et al. (2006) have achieved resolving powers of 50 with tetraheptyl-ammonium  
ions in a peculiarly shaped (isopotential) DMA where the area ratio between the laminarization screen  
and the sampling location of classified particles was only 1.27. The high performance of their analyzer  
persisted up to the maximal flow rate tested of 2300 L/min! In a later study with an improved geometry,  
Martinez-Lozano and Labowsky (2009) achieved a resolving power of 75 with the  
255 | tetraheptylammonium ion.

## 2.6 Is the sheath gas circuit really so complex?

We now return to the issue of other heavy and bulky elements characteristic of a number of previously  
used HRDMAs. Please, note that the exotic applications of DMAs we have pursued at Yale, are not the  
same thing as creating a substantially improved instrument for broad use by aerosol scientists. As an  
260 | example of the potential simplicity of the required system, we note that the large vacuum cleaner  
blowers consuming 1 kW of power we have often used to drive the sheath gas have tended to be large  
inexpensive and relatively inefficient devices, both aerodynamically and electrically. This means that  
these pumps inject a lot of waste heat into the circulating gas, which must be removed by a relatively  
large heat exchanger. However, the recent development of battery operated vacuum cleaners has  
265 | resulted in small blowers with fairly high aerodynamic and electrical efficiencies. Pérez-Lorenzo et al.  
(2017) have described one such commercial pump driving 200 L/min of sheath gas through a HRDMA  
of relatively narrow cross section (inner and outer radii of 4 and 7 mm), while consuming only 12.5

Deleted: Given the need for

Deleted: analyzing

270 Watt. Besides their small dimensions and weight, these efficient blowers heat minimally the  
recirculating gas, making the usual bulky heat exchanger unnecessary.

## 2.7 The role of geometry

Early DMAs were long (small  $k$ ) to favor the classification of large particles. Kousaka et al. (1986)  
demonstrated the interest of short DMAs (larger  $k$ ) to diminish diffusion broadening of ultrafine  
275 particles. Geometry hence enables improving either  $Z_{max}$  or  $Z_{min}$  in plain DMAs by substantial factors.  
There is nevertheless a limit on how small a particle may be analyzed with fair resolving power by  
increasing  $k$  (Rosell et al, 1996). As a result, two or three DMAs of different lengths are commonly  
offered commercially to span a wider size range than coverable by just one DMA. On the other hand, a  
single HFDMA may sweep in a single scan from 1.5 nm to 300 nm, with a resolution in excess of 10.  
280 This is the natural consequence of Equation (1), where a wide change in  $Q$  is equivalent to a wide  
change in  $k$ .  
Our discussion has considered mainly traditional DMA geometries, involving cylindrical or slightly  
converging DMAs. The reason for this narrow scope is that these *axial flow* configurations are the only  
ones where relatively high resolution and flow rate have been demonstrated to date. Flagan and his  
285 students have developed and demonstrated the advantages of so-called radial flow DMAs (Zhang et al.  
1995). Radial and axial DMAs may be designed in a rich range of configurations far beyond what has  
been tested to date. Those so far explored generally have a flow field more or less perpendicular to the  
electric field. Diffusive broadening arises mainly orthogonally to the fluid streamlines, resulting in a  
resolving power scaling as  $V^{1/2}$  [Equation (3)]. It then appears with considerable generality that the  
290 optimal resolving power at given  $Q$  is reached at a certain optimal DMA length and cannot be further  
increased by geometrical manipulations (Fernandez de la Mora, 2002). Unless some unsuspected  
scheme is discovered, achieving high resolving power with ultrafine particles in these geometries  
necessarily requires relatively high flow rates. However, the situation is far more favorable when the  
electric and the flow field are approximately opposed to each other, in which case the resolving power  
295 scales as  $V$  instead of  $V^{1/2}$ . This may be theoretically demonstrated in a one-directional flow, for  
instance created between two porous plates held at different potentials. There is in principle no need of  
sheath gas, though the classification is not differential but cumulative, and a means to avoid complete  
loss of the particles through the porous medium is required. Differential separation can nevertheless be  
achieved by combining opposing axial electric and flow fields with smaller lateral fields, as first  
300 demonstrated by the *Drift-DMA* configuration of Loscertales (1998). This most original proposal  
remained purely conceptual until Tammet (2011) implemented and tested it based on an inclined grid.  
The Opposed Migration Aerosol Classifier of Flagan (2004) adopted the one-dimensional geometry  
with two planar porous surfaces passing sheath gas rather than the aerosol (which moved laterally).  
Neither of these configurations has yet demonstrated high resolving power, but their clear conceptual  
305 advantages suggest that it should be possible to create differential or cumulative mobility analyzers  
achieving high resolving power without requiring unusually large flow rates. Labowsky's isopotential  
DMAs (Martinez Lozano et al. 2006, 2009) are interesting cases where the flow and electric fields in  
the vicinity of the axis are opposed to each other, creating a stagnation point for the particle trajectories.  
This stagnation may viewed as locally analogous to what happens globally in the strictly one



310 dimensional uniform opposing field configuration. The isopotential DMA has demonstrated high  
resolving power, but not at modest flow rates.

### **2.8 The use of high flow DMAs in monodisperse particle production**

315 The first anonymous referee has noted that further discussion of the aerosol sample flow rate is of clear  
broad interest. This is nicely illustrated by the pioneering work of Hontañón and Kruijs, whose study  
achieved the classification of substantial flow rates of monodisperse nanoparticles through the  
development of a DMA of heroic dimensions. Their goal was substantially different from that of the  
Opinion under discussion. Nevertheless, the referee's remark brings the important point that a small  
320 hand-held DMA capable of high flow rate of sheath gas should also be able to handle high  
monodisperse aerosol flow rates, opening up another conceivable application of the DMA development  
I was trying to encourage. This interesting application would perhaps require somewhat different  
geometries (wider inlet and outlet slits, etc.). The main lesson following from the referee's insightful  
remark is that the flow rates may be increased by expanding the DMA dimensions while still running at  
moderate Reynolds numbers, as in the critical DMA developed by Hontañón and Kruijs. The same goal  
325 may be reached by maintaining conventional moderate dimensions while increasing the Reynolds  
number well above 2000. This situation is not incompatible with laminar operation in properly designed  
DMAs, where steady flow has been demonstrated many times with Reynolds numbers well beyond  
100,000.

### **3. Technical problems requiring further developments**

330 **3.1 Detection.** We have already mentioned Hering's key point that high flow CPCs do not presently  
exist and would need to be developed to exploit more fully the high sensitivity potential of HFDMA. A  
recent study by Stolzenburg et al. (2023) has already operated a high flow Reischl DMA with a CPC  
drawing up to 2.5 l/min and demonstrated drastic sensitivity improvements in new particle formation  
335 research.

The discussions from Gerhard Steiner and one anonymous referee have perceptively noted the possible  
advantages of electrical detectors, whose flow rates can be increased with far greater flexibility than in  
CPCs. Many readers may be skeptical about electrical detectors in atmospheric measurements, given  
that the typical electrometers used in aerosol research have noise levels of about 1-10 fA and response  
340 times of 1 s or more. They then require over  $10^3$  elementary charges/s (or particles/s). Nevertheless,  
there is ample room for improving the response time and the sensitivity of most commercial aerosol  
electrometers. We recently reported the incorporation into a collecting filter of an operational amplifier  
circuit developed by Heinz Burtscher and his colleagues, which achieved a response time below 100 ms  
and a noise level of about 0.1 fA (Fernandez de la Mora et al. 2017). Nevertheless, this improvement is  
345 still modest compared to what could be achieved in practice. For instance, CCD cameras typically  
convert individual photons into elementary charges, and then measure their current. Such rapidly  
evolving detectors could therefore be applied to aerosol sensing. That the single particle sensitivity of  
CPCs is in principle extendible to electrical measurements has been clear for some time. For instance,  
Ma et. al. (2017) state that "... the 1.1  $\mu\text{m}$  pixel-pitch device achieves  $0.21e^-$  rms average read noise  
350 with average dark count rate per pixel less than  $0.2e^-/s$ , and 1040 fps readout rate". Notice that the  
outstanding sensitivity of their CMOS-based photon-counting image sensor applies not just to a single

355 detector, but to arrays of many detectors ordered in two dimensions. This means that DMAs with such  
360 detector arrays would not need to scan over the voltage to obtain a size spectrum. The whole spectrum  
365 would be recorded almost instantly in a multitude of detectors distributed along the inner electrode. This  
370 possibility has already been demonstrated (Perez-Lorenzo et al. 2020) in a planar DMA with 100  
operational amplifiers, though not yet with single charge sensitivity. Their work shows that the  
insulating steps separating the various metallic collectors do not cause turbulent transition. A  
commercial linear detector array based on CCD technology with a noise level of 0.5 fA has existed for  
some time (<https://www.jas-sg.com/ids-2030-charged-particle-detector.html>). It is 51 mm long, has  
2126 active pixels each 21  $\mu\text{m}$  wide and 1500  $\mu\text{m}$  high. It has been used in mass spectrometry based on  
instruments that disperse the ions in space. As far as we know, it has not been tested for mobility  
measurements, so, whether or not the collector roughness would cause turbulent transition is unknown.  
An array of electrical detectors with single charge sensitivity would be clearly superior to a DMA  
connected to a single CPC by a large factor equal to the number of mobility channels used. Note that the  
exceptional low noise achieved by Fossum and colleagues relies on the small area of each pixel. A  
typical situation with 1 million pixels, each 1.1 micron in length would appear at first sight to divide  
the signal into too many mobility channels, each receiving too little signal to be measurable in typical  
aerosol applications. Nevertheless, if the claimed ability by Ma et al (2017) of achieving single photon  
counting is correct, the signal would not be diluted over a wide mobility range, but assigned to precise  
values of the mobility. High mobility resolution would not imply in this case reduced sensitivity. One  
could certainly add the counts from as many pixels as desired by post-processing, without loss of the  
available high resolution information. Of course, the fact that such developments are in principle  
possible does not necessarily mean that they will be available soon.  
The point of the first referee that the high sample flow rates manageable by electrometers may  
sometimes achieve greater sensitivities than existing CPCs is indeed nicely illustrated in various  
instruments developed by Tammet and his colleagues at Tartu (See for instance Asmi et al. 2009). It  
would be difficult to match the vast sample flow rates of these specialized instruments in a DMA  
geometry. Nevertheless, it is conceivable that new instruments with resolving powers and sample flow  
rates intermediate between those of the Tartu analyzers and high flow DMAs could be developed by  
imaginative researchers.

### **3.2 DMA Characterization**

385 Until recently, most existing HRDMAs had been characterized with respect to resolution at relatively  
high  $Q/q$ , but not under other flow rate conditions more common in many aerosol studies. Information  
on aerosol losses was also lacking, complicating the inversion of measured mobility spectra. This  
problem has fortunately been redressed in the recent study of

### **3.3 Sheath flow rate control**

390 Anonymous referee 2 brings up several important issues relating to the accurate control of the sheath gas.  
The proposed sweeping over  $Q$  rather than  $V$  requires a continuous way to measure  $Q$ , which cannot be  
achieved with the same precision and ease as a voltage measurement. There is also the issue of the slower  
response time of mechanical pumps versus high voltage sources. Nevertheless, an accurate mobility scale  
requires a precise determination of  $Q$ , irrespective of whether one sweeps over  $Q$  or over  $V$ . Whatever

395 flowmeter is used, the ambiguity in the mobility scale will typically be determined by the flowmeter error,  
which will not be necessarily greater in a  $Q$ -sweep than in a  $V$ -sweep. The main difference is that in a  $V$ -  
sweep the  $Z$  scale will be off by a fixed unknown factor, versus a variable equally unknown factor in a  $Q$ -  
sweep. Assuming that the relative error in the flow rate is not greater at high  $Q$  than at low  $Q$ , it would be  
of little consolation to know that this error is constant. And this slight advantage will be lost in a  
400 conventional sweep if one decides not to be absolutely limited by operating always at fixed  $Q$ , irrespective  
of the size range of interest. These considerations naturally do not remove the need to measure  $Q$  as  
precisely as possible over the whole range of the  $Q$ -sweep, which will evidently require a certain  
development. We are familiar with commercial flowmeters claiming 2% error in  $Q$ , and reproducibility  
within 1%, though only covering the limited range 0-300 L/min. Venturi flowmeters, for instance, can go  
405 substantially higher, with comparable precision. These various flowmeters have response times much  
faster than the pump (limited by the inertia of the rotor), being capable of determining the actual flow rate  
at any instant during a sweep. The instantaneous pump frequency is also readily measured, and is an  
alternative marker of the flow rate. There should accordingly be no difficulty in carrying out minute long  
full  $Q$  and  $V$  scans. The possibility to achieve wide  $V$ -scans in a few seconds has been previously  
410 demonstrated (Fernandez de la Mora et al. 2017b) with electrometers, and would be far more problematic  
on a  $Q$ -sweep.

410 There are, however, other potential difficulties associated to a  $Q$ -scan that may not be so simple to handle,  
or that could limit the scan speed. One of them is that the working pressure in a typical closed sheath flow  
circuit may depend on the pump speed, while the response time for this pressure depends on how the  
system pressure is set relative to the external world. No experience is available at present to assess this  
415 issue.

415 The ability to control the chemical composition of the sheath gas is relatively limited, as it is unpractical  
to dry a large flow rate of sheath gas. The most common way of operating a HRDMA is with the sheath  
gas in closed circuit, such that, after an initial transient, its composition matches that of the entering  
aerosol. Zinola et al. (2019) have operated a high temperature Half-Mini DMA that sampled hot  
420 automobile exhaust gases from the ambient. To avoid vapor condensation from the exhaust gases, they  
did run the sheath gas at 200° C. This was achieved by taking filtered ambient gas into the pump, blowing  
it into a heater, then into the DMA sheath gas circuit inlet, and finally exhausting the excess gas back to  
the atmosphere. The temperature was thus controlled, but not the humidity. Note that the pressure at the  
425 polydisperse aerosol inlet slit in the Half-Mini DMA is close to the pressure in the separation region,  
which is below the pressure at the entry of the sheath gas because the flow moves at relatively high speed  
(Venturi effect). This pressure is also below the sheath outlet pressure because the DMA is provided with  
a diffuser. Accordingly, given a suitable diffuser, it is possible to suck ambient aerosol when the sheath  
gas is exhausting to the atmosphere. On the other hand, none of the HRDMAs built to date to classify  
430 particles as large as 60 nm or more has incorporated a diffuser.

430 In the two previously tested modes, with open and closed sheath gas circuits, the only available  
compositions are either that of the ambient air or that of the aerosol. If one needs a sheath gas drier than  
the aerosol, one option is to dry the aerosol and operate in close circuit. Another option is to recirculate  
the sheath flow through a drying medium. It would be far simpler to remove the little humidity brought by  
435 the aerosol into the closed circuit than to thoroughly dry a large flow of ambient air. Other humidity  
controls are feasible. For instance, if the polydisperse aerosol flow is 3 L/min, and one introduces 27

L/min of dry gas into the closed circuit and (after thorough mixing of these two inputs) draws another 27 L/min out of the circuit, the sheath gas will reach an equilibrium humidity 1/10<sup>th</sup> that of the entering aerosol.

#### 440 **4. Conclusions**

I hope the general considerations and concrete examples provided here will help dispel the notion that HFDMAs are unsuited for general aerosol studies, perhaps also stimulating their commercial development together with that of fast high-flow CPCs. The benefits to aerosol research would be very worth the effort.

#### 445 **Competing interests**

The author is involved in the commercialization of several HRDMAS

#### **Acknowledgments:**

I am much indebted to Susanne V. Hering and Michel Attoui for their insightful remarks on a first draft of this manuscript, and to the editors of Aerosol Research for their invitation to submit this study.

450 Funding from AFOSR grant FA9550-22-1-0097 is gratefully acknowledged.

#### **References**

- Asmi, E., Sipilä, M., Manninen, H. E., Vanhanen, J., Lehtipalo, K., Gagné, S., Neitola, K., Mirme, A., Mirme, S., Tamm, E., Uin, J., Komsaare, K., Attoui, M., and Kulmala, M. 2009, Results of the first air ion spectrometer calibration and intercomparison workshop, *Atmos. Chem. Phys.*, 9, 141–154, <https://doi.org/10.5194/acp-9-141-2009>.
- 455 <https://doi.org/10.5194/acp-9-141-2009>.
- Bezantakos, S., Giamarelou, M., Huang, L., Olfert, J. and Biskos, G.: 2016. Modification of the TSI 3081 differential mobility analyzer to include three monodisperse outlets: Comparison between experimental and theoretical performance, *Aerosol Science and Technology*, 50:12, 1342-1351
- 460 [Cai, R., Chen, D. R., Hao, J., & Jiang, J. \(2017a\). A miniature cylindrical differential mobility analyzer for sub-3 nm particle sizing. \*Journal of Aerosol Science\*, 106, 111–119.](#)
- de Juan, L., Fernandez de la Mora, J.: 1998, High resolution size analysis of nanoparticles and ions: running a Vienna DMA of near optimal length at Reynolds numbers up to 5000, *J. Aerosol Science*, 29, 617-626.
- Eichler, T., de Juan, L. and Fernández de la Mora, J.: 1998. Improvement of the resolution of TSI's 3071 DMA via redesigned sheath air and aerosol inlets, *Aerosol Sci. & Tech.*, 29, 39-49.
- 465 Fernandez de la Mora, J.: 2017. Expanded flow rate range of high-resolution nanoDMAs via improved sample flow injection at the aerosol inlet slit, *J. Aerosol Sci.* 113, 265-275

- 470 Fernandez de la Mora, J., Perez-Lorenzo, L.J., Arranz, G., Amo-Gonzalez, M. and Burtscher, H.: 2017a. Fast high-resolution nanoDMA measurements with a 25 ms response time electrometer, *Aerosol Science and Tech.*, 51(6), 724 – 734
- 475 Fernandez de la Mora, J., Gao, P. and Perez-Lorenzo, L. J.:2017b. Eliminating the main source of instability and turbulence in TSI's 3071 DMA, *Aerosol Sci. & Tech.*, 51(7), 896–902  
[Fernandez de la Mora, J., Angel, D.M., Peccia, J., 2021, How narrow is the gas phase mobility distribution of enveloped viruses? The case of the  \$\Phi 6\$  bacteriophage. \*Anal. Chem.\* 93, 38, 12938–12943](#)
- 480 Fernandez de la Mora, J., Kramar, J. Farkas, N., Papanu, S., Dana, R.; 2023. Singularly narrow size distributions of Colloidal Metrics 200 nm polystyrene latex spheres determined by high resolution mobility analysis. *J. Aerosol Sci.* 170, 106158
- 485 Fernandez de la Mora, J., Papanu, S.; 2023. Characterization of polystyrene latex spheres 300 nm in diameter with singularly narrow size distributions, *J. Aerosol Sci.*, xxx (xxxx) 106230
- 490 Fernandez de la Mora, L.J. Perez-Lorenzo, G. Arranz, M. Amo-Gonzalez, H. Burtscher, Fast high-resolution nanoDMA measurements with a 25 ms response time electrometer, *Aerosol Science and Tech.*, 51(6), 724 – 734, 2017
- 495 Flagan, R. C.; 1998. History of Electrical Aerosol Measurements, *Aerosol Science and Technology*, 28:4, 301-380
- 500 Flagan, R.C.; 1999. On Differential Mobility Analyzer Resolution, *Aerosol Sci. & Tech.*, 30:6, 556-570, Flagan, R.C.; 2004. Opposed Migration Aerosol Classifier (OMAC), *Aerosol Sci. Technol.* 38:890–899.  
[Hontañón, E., Kruis, F. E., 2009, A Differential Mobility Analyzer \(DMA\) for Size Selection of Nanoparticles at High Flow Rates, \*Aerosol Sci. & Techn.\*, 43:1, 25-37, <https://doi.org/10.1080/02786820802446812>](#)
- 505 [Kangasluoma, J., Ahonen, L. R., Laurila, T. M., Cai, R., Enroth, J., Mazon, S. B., Korhonen, F., Aalto, P. P., Kulmala, M., Attoui, M., Petäjä, T., 2018, Laboratory verification of a new high flow differential mobility particle sizer, and field measurements in Hyytiälä. \*J. Aerosol Sci.\* 124, 1-9](#)
- 510 Keck, L., Spielvogel, J. and Grimm, H.; 2008. From NANO particles to large aerosols-fast measurements methods for size and concentration. *Nanosafe 2008: Journal of Physics: Conference Series* 170 (2009) 012007 doi:10.1088/1742-6596/170/1/012007.  
<https://iopscience.iop.org/article/10.1088/1742-6596/170/1/012007/pdf>; accessed 5/July/2023.
- 515 Knutson, E. O. and Whitby, K. T.: 1975. Aerosol classification by electric mobility: Apparatus theory and applications. *Journal of Aerosol Science*, 6, 443–451
- 520 Kong, W., Amanatidis, S., Mai, H., Kim, C., Schulze, B. C., Huang, Y., Lewis, G. S., Hering, S. V., Seinfeld, J. H. and Flagan, R. C.: 2021. The nano-scanning electrical mobility spectrometer (nSEMS) and its application to size distribution measurements of 1.5–25 nm particles, *Atmospheric Measurement Techniques*, 14, 5429–5445
- 525 Kousaka, Y., Okuyama, K. and Mimura, T.: 1986. Effect of Brownian diffusion on electrical classification of ultrafine particles in differential mobility analyzers. *J. Chem. Eny. Japan* 19, 401-407.
- 530 Lee, H., Lee, H. K., Ahn, K.H.: 2020. Numerical simulation of a multi-port differential mobility analyzer (MP-DMA) used in a real-time nano-particle sizer (NPS), *Powder Technology* 362, 135–144
- 535 Liu, B.Y.H., and Pui, D.Y.H.: 1974. A Submicron Aerosol Standard and the Primary Absolute Calibration of the Condensation Nucleus Counter. *J. Colloid Interface Sci.* 47:155-171.

- 510 Loscertales, I.G.: 1998. Drift differential mobility analyzer *Journal of Aerosol Science*, 29(9), 1117-1139
- [Ma, S. Masoodian, D. A. Starkey, E. R. Fossum, 2017, Photon-number-resolving megapixel image sensor at room temperature without avalanche gain, \*Optica\*, 4 \(12\) 1474-1481](https://doi.org/10.1364/OPTICA.4.001474)
- 515 Martínez-Lozano, P., Labowsky, M. and Fernandez de la Mora, J.: 2006. Experimental tests of a nano-DMA with no voltage change between aerosol inlet and outlet slits, *J. Aerosol Science*, 37, 1629-1642
- Martínez-Lozano, P. and M. Labowsky: 2009. An experimental and numerical study of a miniature high resolution isopotential DMA, *J. Aerosol Sci.*, 40(5), 451-462
- 520 Pérez-Lorenzo, L.J. and Fernandez de la Mora, J.: 2017. Powering a portable supercritical DMA with a small battery-operated vacuum cleaner blower, *Aerosol Sci. & Technol.*, 104, 100-105.
- Perez Lorenzo, L.J., O'Mahony, R., Amo-Gonzalez, M. and Fernandez de la Mora, J.: 2020a. Instant Acquisition of High Resolution Mobility Spectra in a Differential Mobility Analyzer with 100 Independent Ion Collectors: Instrument calibration, *Aerosol Sci. & Technol.* 54(10) 1144-1156
- 525 Perez-Lorenzo, L. J., Khanna, V., Meena, T. and Fernandez de la Mora, J.: 2020b. A high resolution DMA covering the 1-67 nm size range, *Aerosol Sci & Technol.*, 54:1, 128-142
- Perez-Lorenzo, L. J. and Fernandez de la Mora, J.: 2021. Exceeding a resolving power of 50 for virus size determination by Differential Mobility Analysis. *J. Aerosol Science*, [151](https://doi.org/10.1016/j.jaerosci.2021.105658), 105658.
- Rosell, J., Loscertales, I. G., Bingham, D. and Fernández de la Mora, J.: 1996. Sizing nanoparticles and ions with a short differential mobility analyzer, *J. Aerosol Science*, 27, 695-719.
- 530 Steiner, G., Attoui, M., Wimmer, D. and Reischl, G. P.: 2010. A Medium Flow, High-Resolution Vienna DMA Running in Recirculating Mode, *Aerosol Science & Technology*, 44:308-315
- Tammet, H.: 2011. Symmetric Inclined Grid Mobility Analyzer for the Measurement of Charged Clusters and Fine Nanoparticles in Atmospheric Air *Aerosol Science & Technology*, 45(4) 468-479
- 535 Winklmayr, W.; Reischl, G.P.; Lindner, A.O.; and Berner, A. (1991) A new electromobility spectrometer for the measurement of aerosol size distributions in the size range from 1 to 1000 nm. *J. Aerosol Sci.* 22:289-29.
- Zhang, S.H., Akutsu, Y., Russell, L. M., Flagan, R. C. and Seinfeld, J. H.: 1995. Radial Differential Mobility Analyzer, *Aerosol Science & Technology*, 23(3), 357-372
- [Zinola, S., Leblanc, M., Rouleau, L., Dunand, X., Baltzopoulou, P., Chasapidis, L., Deloglou, D., Melas, A.D., Konstandopoulos, A. K., Rüggeberg, T., Fierz, M., Burtscher, H., Tejero, A., Amo, M., Zamora, D., 2019, Measurement of Sub-23 nm particles emitted by gasoline direct injection engine with new advanced instrumentation." SAE Technical Paper 2019-01-2195, 2019,](https://doi.org/10.4271/2019-01-2195)

Formatted: Font: Not Bold

Deleted: ¶

I thank the author of RCI: 'Comment on ar-2023-7', Anonymous Referee #1, 29 Oct 2023 for these excellent remarks and bibliographical references. Most of them will be included in the revised manuscript. Notice however that the work of Steiner, Attoui, Wimmer & Reischl was already cited.¶

I agree that further discussion of the aerosol sample flow rate is of clear broad interest, and that the reference to the work of Hontañón and Kruis is most appropriate. Their study achieved the classification of substantial flow rates of monodisperse nanoparticles through the development of a DMA of heroic dimensions. Their goal was substantially different from that of the Opinion under discussion. Nevertheless, the referee's remark brings the important point that a small hand-held DMA capable of high flow rate of sheath gas should also be able to handle high monodisperse aerosol flow rates, opening up another conceivable application of the DMA development I was trying to encourage. This interesting application would perhaps require somewhat different geometries (wider inlet and outlet slits, etc). The main lesson following from the referee's insightful remark is that the flow rates may be increased by expanding the DMA dimensions while still running at moderate Reynolds numbers, as done by Hontañón and Kruis. The same goal may be reached by maintaining conventional moderate dimensions while increasing the Reynolds number well above 2000. This situation is not incompatible with laminar operation in properly designed DMAs, where steady flow has been demonstrated many times with Reynolds numbers well beyond 100,000.¶

The referee reminds us also of the potential interest of using an electrometer as a detector. The prior comment by Dr. Gerhard Steiner has also noted this fact. That the high sample flow rates manageable by electrometers may sometimes achieve greater sensitivities than existing CPCs is nicely illustrated in various Instruments developed by Tammet and his colleagues at Tartu (such as the referee's first reference to Asmi et al.). It would be difficult to match the vast sample flow rates of these specialized instruments in a DMA geometry. Nevertheless, it is clear that this opinion should include a related discussion aimed at stimulating the creation of new instruments with resolving powers and sample flow rates intermediate between those of the Tartu analyzers and DMAs.¶

The referee's reference to Brunelli et al. involves an instrument operating at sheath gas flow rates and resolving powers comparable to those of conventional DMAs. I fully agree with the proposition that the creative geometrical innovations introduced by Professor Flagan and his students in electrostatic classifiers (and their possible future variants) offer great promise for high resolution classification, perhaps even without a need for high flow rates. Nevertheless, it seems to me that this important area of future developments is already discussed in my section 2.7, on the role of geometry.¶

Thank you again!¶

Juan Fernandez de la Mora¶

Dr. Steiner brings up the very important subject of the potential of using electrical detectors as alternatives to CPCs. This view is stated in general, not just in relation to high flow DMAs, though, in our application, electrometers present the obvious advantage of readily accommodating large flow rate variations.¶

Many readers may be skeptical about electrical detectors in atmospheric measurements, given that the typical electrometers used in aerosol research have noise levels of about 1-10 fA and response times of 1 s or more. They then require over 10<sup>3</sup> elementary charges/s (or particles/s). Nevertheless, there is ample room for ¶... [1]

